



PRODUCT SPECIFICATIONS

Incontinence Sensor – P/N IS-XX20-PAD

1. CONSTRUCTION

- 1.A. Length of Switch: 21.75 inches
- 1.B. Width of Switch: 15.75 inches
- 1.C. Cable: 4-conductor, 28AWG, PVC sheathed cable, RJ-9 connector.

2. OPERATION

- 2.A. Activation Req.: Minimum activation: damp area larger than 1” in diameter.
- 2.B. Interface: Normally closed switch. (Other options available upon request).
- 2.C. Max Operating Current: 50 mA
- 2.D. Max Operating Voltage: 24 Volts

3. INSTALLATION

- 3.A. Place IS-XX20 on center of bed or chair, located under the hips and buttocks.

4. CARE

- 4.A. Cleaning: Wipe down with moistened cloth, and wipe dry. Use household cleaners only.

DO NOT MACHINE WASH!

5. Lifetime:

- 5.A. Patient use: Single patient use only. Please dispose after patient no longer requires it.
- 5.B. Washings: Max of 50 washing cycles for a single patient.

6. DIMENSIONS

